

	Application No.	Applicant(s)
Notice of Allowability	10/698,566	BURK, GARY NEIL
	Examiner	Art Unit
	F. L. Evans	2877
The MAILING DATE of this communication appears on the cover sheet with the correspondence address All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS. This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308. 1. This communication is responsive to the papers filed on October 31, 2003.		
2. The allowed claim(s) is/are <u>1-10</u> .		
 3. Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some* c) None of the: 1. Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No. 3. Copies of the certified copies of the priority documents have been received in this national stage application from the 		
International Bureau (PCT Rule 17.2(a)).		
* Certified copies not received: Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application. THIS THREE-MONTH PERIOD IS NOT EXTENDABLE. 4. A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF		
INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.		
5. CORRECTED DRAWINGS (as "replacement sheets") must be submitted.		
(a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached		
1) 🔲 hereto or 2) 🔲 to Paper No./Mail Date		
(b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date		
Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).		
6. DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.		
Attachment(s)		
1. Notice of References Cited (PTO-892)		ratent Application (PTO-152)
2. Notice of Draftperson's Patent Drawing Review (PTO-948)	6. ☐ Interview Summary Paper No./Mail Dat	te
3. Information Disclosure Statements (PTO-1449 or PTO/SB/0 Paper No./Mail Date		
4. Examiner's Comment Regarding Requirement for Deposit	8. 🛛 Examiner's Stateme	ent of Reasons for Allowance
of Biological Material	9.	

Prior Art

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure.

Burk et al (US 6.960,769 B2) disclose a method of standardizing and calibrating a spectroscopic

measuring device.

claim.

Examiner's Statement of Reasons for Allowance

The following is an examiner's statement of reasons for allowance:

As to independent claim 1, the prior art of record, taken alone or in combination, fails to disclose or render obvious a method of operating a measuring sensor having a radiation source and a spectrographic sensor comprising the steps of operating the radiation source to generate a second light signal having a second spectral power distribution, measuring the second light signal with the spectrographic sensor and generating calibration parameters based on a ratio or light sensed around the wavelength of interest and the reference wavelength, in combination with the rest of the limitations of the

As to independent claim 4, the prior art of record, taken alone or in combination, fails to disclose or render obvious a method for calibrating a measuring sensor having a radiation source and a spectrographic sensor comprising the steps controlling the radiation source to generate a calibration light signal and calibrating the standardized spectrographic sensor based on measurements of the calibration light signal taken by the spectrographic sensor, in combination with the rest of the limitations of the claim.

As to independent claim 7, the prior art of record, taken alone or in combination, fails to disclose or render obvious a method for simulating a reference sample in a measuring sensor having a radiation source and a spectrographic sensor comprising the steps set forth in claim 7.

As to independent claim 10, the prior art of record, taken alone or in combination, fails to disclose

Application/Control Number: 10/698,566 Page 3

Art Unit: 2877

or render obvious a simulated reference sample for a measuring sensor, said sample comprising: a system processor for determining a first function from the ratio of the signals representative of the wavelength of interest and the signals representative of the reference wavelength in response to the first radiation, the first ratio function being used to standardize the measuring sensor and for determining a second function from the ratio of the signals representative of the wavelength of interest and the signals representative of the reference wavelength in response to the second radiation, the second ratio function being used to

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

calibrate the measuring sensor, in combination with the rest of the limitations of the claim.

Fax/Telephone Numbers

Any inquiry concerning this communication or earlier communications from the examiner should be directed to the examiner whose telephone number is (571) 272-2414.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Gregory J. Toatley, Jr. can be reached on (571) 272-2800 ext 77. The fax phone number for the organization where this application or proceeding is assigned is (571) 273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

F. L. EVANS PRIMARY EXAMINER ART UNIT 2877